

Calibration Certificate

Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-079	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: BILANCIAI DD700 Capacity: 8000kg Tolerance: +/- 3 Divisions		Works ID Number: Serial Number: T184600558 Model: Increments: 2kg Weighing Range: 8000kg	
--	--	---	--

Repeatability Test			
1st	8000kg	2nd	8000kg
3rd	8000kg		

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
1000kg	1000kg	1000kg
2000kg	2000kg	2000kg
4000kg	4000kg	4000kg
6000kg	6000kg	6000kg
8000kg	8000kg	8000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Off Centre Test - Load		
Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	5340kg	5340kg
Corner/End 1	5340kg	5340kg
Corner/End 2	5340kg	5340kg
Corner 3	5340kg	5340kg
Corner 4	5340kg	5340kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Authorising Signature: 
--

Engineer's Notes:
Relevant Procedure Nos:
Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-081	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: BILANCIAI D800 Capacity: 50000kg Tolerance: +/- 3 Divisions	Works ID Number: Serial Number: 112800841 Model: Increments: 10kg Weighing Range: 44400kg
--	--

Repeatability Test			
1st	44400kg	2nd	44400kg
3rd	44400kg		

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
4000kg	4000kg	4000kg
8000kg	8000kg	8000kg
12000kg	12000kg	12000kg
16000kg	16000kg	16000kg
20000kg	20000kg	20000kg
29160kg	29160kg	29160kg
39160kg	39160kg	39160kg
44400kg	44400kg	44400kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Authorising Signature: 
--

Off Centre Test - Load		
Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	9340kg	9340kg
Corner/End 1	9340kg	9340kg
Corner/End 2	9340kg	9340kg
Corner 3	93340kg	93340kg
Corner 4	9340kg	9340kg
Corner 5	9340kg	9340kg
Corner 6	9320kg	9320kg
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Engineer's Notes:
Relevant Procedure Nos:
Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer:

Seaforde Scrap Metals
50 Farranfad Road
Seaforde
Downpatrick
Down
BT30 8NH

Certificate Number

CC-SEA003001-080

Invoice/Job Number

Test Date

10 January 2024

Next Test Due

January 2025

Weight Set (Cert. No.)

GL01 (T367158)

Date Set Tested

October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details:

Location: Site
Make: BILANCIAI D800
Capacity: 3000kg
Tolerance: +/- 3 Divisions

Works ID Number:

Serial Number: 153505

Model:

Increments: 1kg

Weighing Range: 3000kg

Repeatability Test

1st 2nd 3rd

Linearity Test

Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
600kg	600kg	600kg
1200kg	1200kg	1200kg
1800kg	1800kg	1800kg
2400kg	2400kg	2400kg
3000kg	3000kg	3000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Off Centre Test - Load

Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	600kg	600kg
Corner/End 1	600kg	600kg
Corner/End 2	600kg	600kg
Corner 3	600kg	600kg
Corner 4	600kg	600kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Authorising Signature:



Engineer's Notes:

Relevant Procedure Nos:

Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
Mullavilly, Co. Armagh, BT62 2JX
Tel: 028 3884 0714
Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
Farmers Cross, Co. Cork T12 T854
Tel: 021 431 8388
Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer:

Seaforde Scrap Metals
50 Farranfad Road
Seaforde
Downpatrick
Down
BT30 8NH

Certificate Number

CC-SEA003001-083

Invoice/Job Number

Test Date

10 January 2024

Next Test Due

January 2025

Weight Set (Cert. No.)

GL01 (T367158)

Date Set Tested

October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details:

Location: Site
Make: BILANCIAI D800
Capacity: 3000kg
Tolerance: +/- 3 Divisions

Works ID Number:

Serial Number: T070100231

Model:

Increments: 1kg

Weighing Range: 3000kg

Repeatability Test

1st 2nd 3rd

Linearity Test

Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
600kg	600kg	600kg
1200kg	1200kg	1200kg
1800kg	1800kg	1800kg
2400kg	2400kg	2400kg
3000kg	3000kg	3000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Off Centre Test - Load

Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	600kg	600kg
Corner/End 1	600kg	600kg
Corner/End 2	600kg	600kg
Corner 3	600kg	600kg
Corner 4	600kg	600kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Authorising Signature:

Engineer's Notes:

Relevant Procedure Nos:

Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
Mullavilly, Co. Armagh, BT62 2JX
Tel: 028 3884 0714
Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
Farmers Cross, Co. Cork T12 T854
Tel: 021 431 8388
Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-082	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: BILANCIAI D800 Capacity: 3000kg Tolerance: +/- 3 Divisions	Works ID Number: Serial Number: 154018 Model: Increments: 1kg Weighing Range: 3000kg
---	---

Repeatability Test		
1st	3000kg	2nd
		3rd
	3000kg	

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
600kg	600kg	600kg
1200kg	1200kg	1200kg
1800kg	1800kg	1800kg
2400kg	2400kg	2400kg
3000kg	3000kg	3000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Off Centre Test - Load		
Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	600kg	600kg
Corner/End 1	600kg	600kg
Corner/End 2	600kg	600kg
Corner 3	600kg	600kg
Corner 4	600kg	600kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Engineer's Notes:
Relevant Procedure Nos:
Engineer's Name: Andrew Gray

Authorising Signature: 
--

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-084	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: BILANCIAI D800 Capacity: 25kg Tolerance: +/- 3 Divisions	Works ID Number: Serial Number: 104100671 Model: Increments: 1kg Weighing Range: 25kg
---	--

Repeatability Test		
1st	25kg	2nd
		25kg
3rd	25kg	

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
5kg	5kg	5kg
10kg	10kg	10kg
15kg	15kg	15kg
20kg	20kg	20kg
25kg	25kg	25kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Authorising Signature: 
--

Off Centre Test - Load		
Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	25kg	25kg
Corner/End 1	25kg	25kg
Corner/End 2	25kg	25kg
Corner 3	25kg	25kg
Corner 4	25kg	25kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Engineer's Notes:
Relevant Procedure Nos:
Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer:

Seaforde Scrap Metals
50 Farranfad Road
Seaforde
Downpatrick
Down
BT30 8NH

Certificate Number

CC-SEA003001-085

Invoice/Job Number

Test Date

10 January 2024

Next Test Due

January 2025

Weight Set (Cert. No.)

GL01 (T367158)

Date Set Tested

October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details:

Location: Site
Make: Bilanciai Flynet 50
Capacity: 50000kg
Tolerance: +/- 3 Divisions

Works ID Number:

Serial Number: T222102132

Model:

Increments: 20kg

Weighing Range: 44400kg

Repeatability Test

1st 44400kg 2nd 44400kg 3rd 44400kg

Linearity Test

Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
4000kg	4000kg	4000kg
8000kg	8000kg	8000kg
12000kg	12000kg	12000kg
16000kg	16000kg	16000kg
20000kg	20000kg	20000kg
29160kg	29160kg	29160kg
39160kg	39160kg	39160kg
44400kg	44400kg	44400kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Off Centre Test - Load

Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	9340kg	9340kg
Corner/End 1	9340kg	9340kg
Corner/End 2	9340kg	9340kg
Corner 3	9320kg	9320kg
Corner 4	9340kg	9340kg
Corner 5	9320kg	9320kg
Corner 6	9320kg	9320kg
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Engineer's Notes:

Relevant Procedure Nos:

Engineer's Name: Andrew Gray

Authorising Signature:



Head Office: 118 Portadown Road,
Mullavilly, Co. Armagh, BT62 2JX
Tel: 028 3884 0714
Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
Farmers Cross, Co. Cork T12 T854
Tel: 021 431 8388
Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate

Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-086	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: Bilanciai Flynet 50 Capacity: 8000kg Tolerance: +/- 3 Divisions	Works ID Number: Serial Number: T210400532 Model: Increments: 2kg Weighing Range: 8000kg
--	---

Repeatability Test		
1st	8000kg	2nd
		8000kg
3rd	8000kg	

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
1000kg	1000kg	1000kg
2000kg	2000kg	2000kg
4000kg	4000kg	4000kg
6000kg	6000kg	6000kg
8000kg	8000kg	8000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Authorising Signature: 
--

Off Centre Test - Load		
Test Point	Reading Before Adjustment	Reading After Adjustment
Centre	7340kg	7340kg
Corner/End 1	7340kg	7340kg
Corner/End 2	7340kg	7340kg
Corner 3	7340kg	7340kg
Corner 4	7340kg	7340kg
Corner 5	N/A	N/A
Corner 6	N/A	N/A
Corner 7	N/A	N/A
Corner 8	N/A	N/A

Engineer's Notes:
Relevant Procedure Nos:
Engineer's Name: Andrew Gray

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426

web: www.globeweigh.com email: sales@globeweigh.com

Calibration Certificate


Customer: Seaforde Scrap Metals 50 Farranfad Road Seaforde Downpatrick Down BT30 8NH	Certificate Number CC-SEA003001-087	Invoice/Job Number
	Test Date 10 January 2024	Next Test Due January 2025
	Weight Set (Cert. No.) GL01 (T367158)	Date Set Tested October 2023

The following certification tests have been carried out using calibrated weights traceable to national standards :

Equipment Details: Location: Site Make: BILANCAI DD700 Capacity: 3000kg Tolerance: +/- 3 Divisions		Works ID Number: Serial Number: T220200478 Model: Increments: 1kg Weighing Range: 3000kg
---	--	---

Repeatability Test		
1st	3000kg	2nd
		3rd

Linearity Test		
Test Point	Reading Before Adjustment	Reading After Adjustment
0kg	0kg	0kg
600kg	600kg	600kg
1200kg	1200kg	1200kg
1800kg	1800kg	1800kg
2400kg	2400kg	2400kg
3000kg	3000kg	3000kg
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Authorising Signature: 
--

Head Office: 118 Portadown Road,
 Mullavilly, Co. Armagh, BT62 2JX
 Tel: 028 3884 0714
 Fax: 028 3884 0420



1113

Cork Office: Unit 3, Airport East Business Park,
 Farmers Cross, Co. Cork T12 T854
 Tel: 021 431 8388
 Fax: 021 480 1426